



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-006690	Instrument Location SMITH COUNTY S.O
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): Annual inspect
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 5/29/26
-------------------------------	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen	Certified Date 05June2026
---	-------------------------------------

Checklist #10 was checked by Anna Ingemansen on 05June2026. All three tests were sent to the Crime Laboratory. This does not affect the installation of the instrument.
 AEI 05June2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006690
Location = SACO 8164.14.00 09/16
05/29/2026 16:56

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890- = |
ABCDEFGHIJKLMNopqrstuvwxyz!@#%&^*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 08425080A1
Standard Cyl #? 11
Standard Expiration? 05/05/2027
Oper No? 132607

Flow Cal. Date: 08/23/2018
Slope 677
Intercept -593298

IR Calibration Date: 08/23/2018

	3um	9um
0th Coef(*100):	-8380	-21663
1st Coef(*100):	265142	130429
2nd Coef(*100):	3933	1429
H2O adj(mg/l*10k):	412	539

***** Printer Test End *****



Operator Signature
JONATHAN WOLD

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006690
Location = SACO 8164.14.00 09/16
05/29/2026 17:15

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:16
02 Std. Gas	0.079	17:16
03 Room Air	0.000	17:17
04 Std. Gas	0.080	17:17
05 Room Air	0.000	17:18
06 Std. Gas	0.079	17:18
07 Room Air	0.000	17:19

Lot No = 08425080A1
Cyl No = 11
Exp Date = 05/05/2027
County = 41 Oper No. = 132607


Operator Signature
JONATHAN WOLD

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006690
Location = SACO 8164.14.00 09/16
05/29/2026 17:27

DRY CAL CHECK

Test	AC	Time
01 Room Air	RFI*	17:27
02 Room Air	0.000	17:28

*Invalid Test
Inhibited - RFI

Lot No = 08425080A1
Cyl No = 11
Exp Date = 05/05/2027
County = 41 Oper No. = 132607



Operator Signature
JONATHAN WOLD

Remarks:

Form 106-I8000