



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL  
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 4199 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
  - Gas Cylinder Yes or No (If Yes, Lot # \_\_\_\_\_ Cyl. # \_\_\_\_\_)
  - Keys Yes or No
  - Power Cord Yes or No
2.  Download Data
3.  Upload Operator File
4.  Current Location Code: TOXL
5.  Battery Check
  - Was the external battery pack replaced? Yes or No
6.  O-Rings
  - Replaced Simulator O-Ring Yes or No
  - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1.  Diagnostics passed and instrument in "Ready" mode
2.  Breath tube heated
3.  Date, time and location code (Level 2,E). Re-set if necessary.  
Time Zone: CS or MDT (Time on test records will be in time zone circled).
4.  Print test (Level 1,P). Sign and attach test record.
5.  Tank monitor (Level 3,D,G).  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi  
Display and Regulator ± 50 psi of each other Yes or No  
Gas tank tare necessary? Yes or No  
If Yes, display readings after tare (Level 3,M,C,G):  
Display: \_\_\_\_\_ psi Regulator: \_\_\_\_\_ psi

C. Tests (Sign and attach test records):

1.  Configure simulator for the following test (Level 1,S).  
Wet Calibration Check - Low AC (Level 1,C)  
Known Value ≤ 0.03 AC: 0.020 AC  
Sim. Ser #: MP5290  
Lot #: 202501E  
Exp. Date: 1/28/27  
 Results ± 0.005 of known AC
2.  Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value  $\geq$  0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results  $\pm$  5% AC of known AC

3.  Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 26-4437

Cylinder #: 114

Exp. Date: 1/29/28

4.  Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1CS9

Exp. Date: None

Display reads "Interferent Detect"

5.  Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6.  RFI Check (CMS Mode)

Display reads "RFI Detect"

7.  Dry Calibration Check (Level 1,C)

Test 1 0.078      Test 4 0.078      Test 7 0.077

Test 2 0.078      Test 5 0.077      Test 8 0.077

Test 3 0.077      Test 6 0.077      Test 9 0.078

Average 0.077

Results  $\pm$  0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: \_\_\_\_\_

If Yes, change location code back to A.4 unless A.4 is TOXL.

[Signature]  
Inspector Signature

04 June 2026  
Date

[Signature]  
Reviewer

08 June 2026  
Date

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000            SN 80-004199  
Location = TOXL                    8164.14.00 09/16  
06/04/2026                            14:10

\*\*\*\*\* Printer Test \*\*\*\*\*

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

abcdefghijklmnopqrstuvwxyz1234567890-=\_|  
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#\$\$%^&\*()\_+?

Current Instrument Setup

Data Entry Mode:            Enabled  
Start Test Sequence:        DABACABA  
Display Prelim Rslt?        Yes  
Display Third Digit?        Yes  
Inhib Printer(Y/N)?        No  
Display Volume?            No  
Disable On Memfull?        Yes  
# of Print Copies?          1  
Select Std (D/W/I)?        Dry  
Standard Value?            0.080  
Standard Lot #?            19825080A2  
Standard Cyl #?            36  
Standard Expiration?        09/05/2027  
Oper No?                    133237

Flow Cal. Date:            06/17/2019  
Slope                        593  
Intercept                   -481610

IR Calibration Date:        04/18/2024  
                                  3um                    9um

-----  
0th Coef(\*100):            -25088                    -28296  
1st Coef(\*100):            269537                    131183  
2nd Coef(\*100):            2308                        1373  
H2O adj(mg/l\*10k):        357                        563

\*\*\*\*\* Printer Test End \*\*\*\*\*



Operator Signature  
ANNA INGEMANSEN

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004199  
Location = TOXL      8164.14.00 09/16  
06/04/2026      14:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:12
02 Std. Sol.	0.018	14:13
03 Room Air	0.000	14:14
04 Std. Sol.	0.018	14:15
05 Room Air	0.000	14:15
06 Std. Sol.	0.018	14:16
07 Room Air	0.000	14:16

08 Sim Temp = 34.0°C

Simul Ser No = MP5290  
Std Sol No = 202501E  
County = 08      Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: Low AC check

0.020AL

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004199  
Location = TOXL      8164.14.00 09/16  
06/04/2026      14:22

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:23
02 Std. Sol.	0.293	14:24
03 Room Air	0.000	14:24
04 Std. Sol.	0.293	14:25
05 Room Air	0.000	14:26
06 Std. Sol.	0.293	14:26
07 Room Air	0.000	14:27

08 Sim Temp = 34.0°C

Simul Ser No = MP3003  
Std Sol No = 202408H  
County = 08      Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks:

High AC Check  
0.300 AC

Form 106-I8000




Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004199  
Location = TOXL                 8164.14.00 09/16  
06/04/2026                        14:38

Test	AC	Time
01 Room Air	0.000	14:39
02 Invalid Sample	X.XXX	14:39
03 Room Air	0.000	14:39

\*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                     Weight = NA  
Test = OTH                         Cit = INVALID  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 114  
Expiration Date = 01/29/2028  
County = 08                        Oper No. = 133237



Operator Signature  
ANNA INGEMANSEN

Remarks: Invalid sample  
              Mouth alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004199  
Location = TOXL                    8164.14.00 09/16  
06/04/2026                            14:41

Test	AC	Time
01 Diagnostic	OK	14:41
02 Room Air	0.000	14:42
03 *Subject Test	RFI*	14:42
04 Room Air	0.000	14:43

\*Invalid Test  
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE  
Sub DOB = 07/25/1998  
Sub Sex = Male                      Weight = NA  
Test = OTH                            Cit = RFI  
Dr. Lic. = ND/TES989643  
Lot No = 264437  
Cyl No = 114  
Expiration Date = 01/29/2028  
County = 08                          Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature  
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004199  
Location = TOXL                    8164.14.00 09/16  
06/04/2026                            14:48

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:49
02 Std. Gas	0.078	14:49
03 Room Air	0.000	14:49
04 Std. Gas	0.078	14:50
05 Room Air	0.000	14:50
06 Std. Gas	0.077	14:51
07 Room Air	0.000	14:51

Lot No = 264437  
Cyl No = 114  
Exp Date = 01/29/2028  
County = 08                            Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks: DRY CAL CHECK  
              #1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer      Alcohol Analyzer  
North Dakota Model 8000      SN 80-004199  
Location = TOXL      8164.14.00 09/16  
06/04/2026      14:52

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:52
02 Std. Gas	0.078	14:52
03 Room Air	0.000	14:53
04 Std. Gas	0.077	14:53
05 Room Air	0.000	14:54
06 Std. Gas	0.077	14:54
07 Room Air	0.000	14:55

Lot No = 264437  
Cyl No = 114  
Exp Date = 01/29/2028  
County = 08      Oper No. = 133237

  
Operator Signature  
ANNA INGEMANSEN

Remarks:

Dry Cal check  
#4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist  
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer            Alcohol Analyzer  
North Dakota Model 8000        SN 80-004199  
Location = TOXL                    8164.14.00 09/16  
06/04/2026                            14:55

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:56
02 Std. Gas	0.077	14:56
03 Room Air	0.000	14:57
04 Std. Gas	0.077	14:57
05 Room Air	0.000	14:57
06 Std. Gas	0.078	14:58
07 Room Air	0.000	14:58

Lot No = 264437  
Cyl No = 114  
Exp Date = 01/29/2028  
County = 08                            Oper No. = 133237

  
\_\_\_\_\_  
Operator Signature  
ANNA INGEMANSEN

Remarks:    Dry cal check  
                  #7-9

Form 106-I8000