



**NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION**

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 666666 Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:
 - Gas Cylinder Yes or No (If Yes, Lot # _____ Cyl. # _____)
 - Keys Yes or No
 - Power Cord Yes or No
2. Download Data
3. Upload Operator File
4. Current Location Code: TOXL
5. Battery Check
 - Was the external battery pack replaced? Yes or No
6. O-Rings
 - Replaced Simulator O-Ring Yes or No
 - Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. Diagnostics passed and instrument in "Ready" mode
2. Breath tube heated
3. Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. Print test (Level 1,P). Sign and attach test record.
5. Tank monitor (Level 3,D,G).
Display: 727 psi Regulator: 725 psi
Display and Regulator ± 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value ≤ 0.03 AC: 0.020 AC
Sim. Ser #: MP529D
Lot #: 202501E
Exp. Date: 1/28/27
 Results ± 0.005 of known AC
2. Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value \geq 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

Results \pm 5% AC of known AC

3. Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 26-4437

Cylinder #: 114

Exp. Date: 1/29/28

4. Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: ICS9

Exp. Date: None

Display reads "Interferent Detect"

5. Invalid Sample (Level 1,B)

Display reads "Invalid Sample"

6. RFI Check (CMS Mode)

Display reads "RFI Detect"

7. Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.081 Test 7 0.080

Test 2 0.081 Test 5 0.080 Test 8 0.080

Test 3 0.081 Test 6 0.081 Test 9 0.082

Average 0.080

Results \pm 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL.

Allyn Lyng
Inspector Signature

04 June 2026
Date

Janelle Pitschler
Reviewer

08 June 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:09

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNPOQRSTUVWXYZ!@#%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 112
Standard Expiration? 01/29/2028
Oper No? 133237

Flow Cal. Date: 05/16/2023
Slope 686
Intercept -761238

IR Calibration Date: 11/20/2025
 3um 9um

0th Coef(*100): -17770 -27082
1st Coef(*100): 265146 136552
2nd Coef(*100): 2379 1305
H2O adj(mg/l*10k): 412 507

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: *print Test*

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:11

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:11
02 Std. Sol.	0.019	08:12
03 Room Air	0.000	08:13
04 Std. Sol.	0.019	08:14
05 Room Air	0.000	08:14
06 Std. Sol.	0.019	08:15
07 Room Air	0.000	08:15

08 Sim Temp = 34.0°C

Simul Ser No = MP5290
Std Sol No = 202501E
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Low AC check
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:16

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:17
02 Std. Sol.	0.297	08:18
03 Room Air	0.000	08:19
04 Std. Sol.	0.297	08:19
05 Room Air	0.000	08:20
06 Std. Sol.	0.297	08:21
07 Room Air	0.000	08:21

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: High AC Check
 0.300Ac

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:23

Test	AC	Time
01 Room Air	0.000	08:24
02 *Subject Test	INT*	08:25
03 Room Air	0.000	08:25

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = ~~DUI~~ Other ac 014126 Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 114
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent check
0.10AC + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:26

Test	AC	Time
01 Room Air	0.000	08:26
02 Invalid Sample	X.XXX	08:27
03 Room Air	0.000	08:28

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 114
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Invalid sample
Mouth Alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:28

Test	AC	Time
01 Diagnostic	OK	08:29
02 Room Air	0.000	08:29
03 *Subject Test	RFI*	08:30
04 Room Air	0.000	08:30

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 114
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237

I followed the Approved Method and the instructions displayed by the Intoxilyzer in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks:

RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:31

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:31
02 Std. Gas	0.080	08:32
03 Room Air	0.000	08:32
04 Std. Gas	0.081	08:33
05 Room Air	0.000	08:33
06 Std. Gas	0.081	08:34
07 Room Air	0.000	08:34

Lot No = 264437
Cyl No = 114
Exp Date = 01/29/2028
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#1-3

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:35
02 Std. Gas	0.081	08:35
03 Room Air	0.000	08:36
04 Std. Gas	0.080	08:36
05 Room Air	0.000	08:37
06 Std. Gas	0.081	08:37
07 Room Air	0.000	08:38

Lot No = 264437
Cyl No = 114
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks:

Dry cal check
#4-6

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006666
Location = TOXL 8164.14.00 09/16
06/04/2026 08:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	08:41
02 Std. Gas	0.080	08:41
03 Room Air	0.000	08:41
04 Std. Gas	0.080	08:42
05 Room Air	0.000	08:42
06 Std. Gas	0.082	08:43
07 Room Air	0.000	08:43

Lot No = 264437
Cyl No = 114
Exp Date = 01/29/2028
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks:

Dry Cal Check
7-9

Form 106-I8000