

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 REPAIR AND MAINTENANCE

Intoxilyzer® 8000 Serial Number: 80-00 4206 Maintenance Location: TOXL

1. Reported Issue and Issue Found:

High calibration values

2. Repair/Maintenance Performed:

Calibration Adjustment

3. Does a Calibration Adjustment need to be performed? Yes or No

- If Yes, perform a calibration check.

Configure dry gas standard (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 20

Exp. Date: 9/5/27

Dry Calibration Check (Level 1,C)

Test 1 0.083 Test 2 0.082 Test 3 0.081

- If No, go to step 4.

4. Does the instrument flow sensor need to be adjusted? Yes or No

- If Yes, complete Flow Sensor Adjustment and Calibration (Document ID: 11872), then go to step 5.
- If No, go to step 5.

5. Does the instrument calibration need to be adjusted? Yes or No
- If Yes, complete Intoxilyzer® 8000 Calibration Adjustment (Document ID: 11859), then go to step 6.
 - If No, go to step 6.
6. Was the instrument cover removed for the repair? Yes or No
- If Yes, complete Intoxilyzer® 8000 Initial Inspection (Document ID: 11784), then go to step 7.
 - If No, go to step 7.
7. Is an Annual Inspection due on the Intoxilyzer® 8000? Yes or No
- If Yes, complete Intoxilyzer® 8000 Annual Inspection (Document ID: 11698), step 8 is not needed, and sign below.
 - If No, go to step 8.
8. Was an Intoxilyzer® 8000 Initial Inspection (Document ID: 11784) completed in step 6? Yes or No
- If Yes, no further action is needed and sign below.
 - If No, Complete Intoxilyzer® 8000 Calibration (Document ID: 11871) and sign below.

Remarks/Notes: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why:


Breath Alcohol Analyst Signature

10 June 2026
Date


Reviewer Signature

10 June 2026
Date


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004206
Location = TOXL 8164.14.00 09/16
05/15/2026 10:37

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:39
02 Std. Gas	0.083	10:39
03 Room Air	0.000	10:39
04 Std. Gas	0.082	10:40
05 Room Air	0.000	10:40
06 Std. Gas	0.081	10:41
07 Room Air	0.000	10:41

Lot No = 19825080A2
Cyl No = 20
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: calibration check
for Repair & Maintenance

Form 106-I8000