



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
 NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
 CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
 SFN 59281 (06/2018)

Serial Number 80-007086	Instrument Location RITCHLAND CO. JAIL, WAHPETON, ND
Reason for Install/Repair <input checked="" type="checkbox"/> Install After Receiving From Crime Laboratory <input type="checkbox"/> Install After Location Change <input type="checkbox"/> Other (Specify) _____	

Check When Done:

- 1. Surge Protector Installed/Property Grounded.
- 2. Telephone Line Connected to Intoxilyzer® 8000.
- 3. Breath Tube Heated.
- 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- 6. Run Tests:
 - A. Print Test (Level 1, Function P).
 - B. ACA Test (Level 1, Function C).
 - C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- 7. Repair and/or Maintenance Performed (if any): N/A
- 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature 	Date 06/11/2026
-------------------------------	---------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen	Certified Date 16 June 2026
---	---------------------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = WAHP 8164.16.00 09/18
06/11/2026 17:44

***** Printer Test *****

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxy1234567890-|=|
ABCDEFGHIJKLMNopQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 19825080A2
Standard Cyl #? 16
Standard Expiration? 09/05/2027
Oper No? 132570

Flow Cal. Date: 10/16/2018
Slope 688
Intercept -653049

IR Calibration Date: 06/04/2025
 3um 9um

0th Coef(*100): -45735 -21942
1st Coef(*100): 240579 133631
2nd Coef(*100): 2470 1307
H2O adj(mg/l*10k): 836 395

***** Printer Test End *****



Operator Signature
CHRISTOPHER HUARD

Remarks: **INSTALL**


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = WAHP 8164.16.00 09/18
06/11/2026 17:45

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:46
02 Std. Gas	0.078	17:46
03 Room Air	0.000	17:47
04 Std. Gas	0.078	17:47
05 Room Air	0.000	17:48
06 Std. Gas	0.077	17:48
07 Room Air	0.000	17:48

Lot No = 19825080A2
Cyl No = 16
Exp Date = 09/05/2027
County = 39 Oper No. = 132570



Operator Signature
CHRISTOPHER HUARD

Remarks: **INSTALL**

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007086
Location = WAHP 8164.16.00 09/18
06/11/2026 17:51

Test	AC	Time
01 Room Air	RFI*	17:53
02 Room Air	0.000	17:54

*Invalid Test
Inhibited - RFI

Sub Name = RFI TEST, NA NA
Sub DOB = 01/01/2026
Sub Sex = Male Weight = 150
Test = DUI Cit = 123456789
Dr. Lic. = ND/1234ND
Lot No = 19825080A2
Cyl No = 16
Expiration Date = 09/05/2027
County = 39 Oper No. = 132570



Operator Signature
CHRISTOPHER HUARD

Remarks: **INSTALL**

Form 106-I8000