



NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6490 Inspection Location: TDXL

A. Pre-Inspection

1. Items with Instrument:
Gas Cylinder Yes or No (If Yes, Lot # 264437 Cyl. # III)
Keys Yes or No
Power Cord Yes or No
2. ☒ Download Data
3. ☒ Upload Operator File
4. ☒ Current Location Code: ASHL
5. ☒ Battery Check
Was the external battery pack replaced? Yes or No
6. ☒ O-Rings
Replaced Simulator O-Ring Yes or No
Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode
2. ☒ Breath tube heated
3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.
Time Zone: CST or MDT (Time on test records will be in time zone circled).
4. ☒ Print test (Level 1,P). Sign and attach test record.
5. ☒ Tank monitor (Level 3,D,G).
Display: 972 psi Regulator: 1000 psi
Display and Regulator $\pm$ 50 psi of each other Yes or No
Gas tank tare necessary? Yes or No
If Yes, display readings after tare (Level 3,M,C,G):
Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).
Wet Calibration Check - Low AC (Level 1,C)
Known Value $\leq$ 0.03 AC: 0.020 AC
Sim. Ser #: MP5290
Lot #: 202501E
Exp. Date: 1/28/27
☒ Results $\pm$ 0.005 of known AC
2. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value $\geq$ 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 1/29/26

☒ Results $\pm$ 5% AC of known AC

3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 19825080A2

Cylinder #: 32

Exp. Date: 9/6/27

4. ☒ Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1659

Exp. Date: None

☒ Display reads "Interferent Detect"

5. ☒ Invalid Sample (Level 1,B)

☒ Display reads "Invalid Sample"

6. ☒ RFI Check (CMS Mode)

☒ Display reads "RFI Detect"

7. ☒ Dry Calibration Check (Level 1,C)

Test 1	<u>0.079</u>	Test 4	<u>0.080</u>	Test 7	<u>0.079</u>
Test 2	<u>0.079</u>	Test 5	<u>0.080</u>	Test 8	<u>0.080</u>
Test 3	<u>0.079</u>	Test 6	<u>0.079</u>	Test 9	<u>0.079</u>
		Average	<u>0.079</u>		

☒ Results $\pm$ 0.005 AC of known AC

D. Remarks/Maintenance: NIA

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

Ann Lymann
Inspector Signature

24 April 2026
Date

Janette Putschler
Reviewer

06 May 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:22

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 111
Standard Expiration? 01/29/2028
Oper No? 133237

Flow Cal. Date: 09/04/2020
Slope 649
Intercept -598817

IR Calibration Date: 09/04/2020

	3um	9um
0th Coef(*100):	-25190	-16392
1st Coef(*100):	271645	136906
2nd Coef(*100):	2533	1222
H2O adj(mg/l*10k):	460	309

***** Printer Test End *****



Operator Signature
ANNA INGEMANSEN

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:24

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:24
02 Std. Sol.	0.021	14:25
03 Room Air	0.000	14:26
04 Std. Sol.	0.021	14:26
05 Room Air	0.000	14:27
06 Std. Sol.	0.020	14:28
07 Room Air	0.000	14:28

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 202501E

County = 08

Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: LOW AlC check
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:30

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:30
02 Std. Sol.	0.294	14:31
03 Room Air	0.000	14:32
04 Std. Sol.	0.295	14:32
05 Room Air	0.000	14:33
06 Std. Sol.	0.295	14:34
07 Room Air	0.000	14:34

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: High AC Check
0.300AC

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:36

Test	AC	Time
01 Room Air	0.000	14:37
02 *Subject Test	INT*	14:38
03 Room Air	0.000	14:38

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 32
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent Check
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:41

Test	AC	Time
01 Room Air	0.000	14:42
02 Invalid Sample	X.XXX	14:42
03 Room Air	0.000	14:43

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 32
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Invalid Sample
Mouth Alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:46

Test	AC	Time
01 Diagnostic	OK	14:47
02 Room Air	0.000	14:47
03 *Subject Test	RFI*	14:49
04 Room Air	0.000	14:49

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI CHECK
Dr. Lic. = ND/TES989643
Lot No = 19825080A2
Cyl No = 32
Expiration Date = 09/05/2027
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI Check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:50
02 Std. Gas	0.079	14:51
03 Room Air	0.000	14:51
04 Std. Gas	0.079	14:52
05 Room Air	0.000	14:52
06 Std. Gas	0.079	14:53
07 Room Air	0.000	14:53

Lot No = 19825080A2
Cyl No = 32
Exp Date = 09/05/2027
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:54

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:54
02 Std. Gas	0.080	14:54
03 Room Air	0.000	14:55
04 Std. Gas	0.080	14:55
05 Room Air	0.000	14:56
06 Std. Gas	0.079	14:56
07 Room Air	0.000	14:57

Lot No = 19825080A2
Cyl No = 32
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#4-b

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006490
Location = TOXL 8164.14.00 09/16
04/24/2026 14:58

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:58
02 Std. Gas	0.079	14:59
03 Room Air	0.000	14:59
04 Std. Gas	0.080	14:59
05 Room Air	0.000	15:00
06 Std. Gas	0.079	15:00
07 Room Air	0.000	15:01

Lot No = 19825080A2
Cyl No = 32
Exp Date = 09/05/2027
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry Cal check
#7-9

Form 106-I8000