



ETHANOL GAS STANDARD CYLINDER REPORT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Cody Baker</i>		
Location <i>515 2nd Ave SW Minot ND (11001)</i>		Intoxilyzer® Serial Number <i>80-006503</i>
Gas Lot Number <i>26-4437</i>	Gas Cylinder Number <i>116</i>	Gas Expiration Date <i>01/29/2028</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.080* AC
B. *0.080* AC
C. *0.080* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>5/22/2026</i>
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen <i>Anna Ingemansen</i>	Date 16 July 2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006503
Location = MIPD 8164.14.00 09/16
05/22/2026 10:49

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:49
02 Std. Gas	0.080	10:50
03 Room Air	0.000	10:50
04 Std. Gas	0.080	10:51
05 Room Air	0.000	10:51
06 Std. Gas	0.080	10:52
07 Room Air	0.000	10:52

Lot No = 264437
Cyl No = 0116
Exp Date = 01/29/2028
County = 64 Oper No. = 133426


Operator Signature
CODY BAKER

Remarks:

Form 106-I8000