

**ETHANOL GAS STANDARD CYLINDER REPORT**

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Travis Collins		
Location BILL		Intoxilyzer® Serial Number 80-005958
Gas Lot Number 26-4437	Gas Cylinder Number 107	Gas Expiration Date 01/29/2028

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
- A. 0.078 AC
B. 0.078 AC
C. 0.078 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☐ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency. NA
- ☒ 5. Send the Following to the Crime Laboratory:
- A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>T Collins</i>	Date 06/23/2026
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen <i>Anna Ingemansen</i>	Date 07Jul2026

Intoxilyzer Test Record and Checklist
ND0AG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer
North Dakota Model 8000
Location = BILL
06/23/2026

Alcohol Analyzer
SN 80-005958
8164.14.00 09/16
09:49

Test	AC	Time
01 Diagnostic	OK	09:53
02 Room Air	0.000	09:53
03 *Subject Test	RFI*	09:53
04 Room Air	0.000	09:54

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST T
Sub DOB = 02/01/2000

Sub Sex = Male

Test = DUI

Dr. Lic. = ND/TE5001234

Lot No = 264437

Cyl No = 107

Expiration Date = 01/29/2028

County = 04

Weight = 238
Cit = 1010010

I followed the Approved Method
and the instructions displayed
by the Intoxilyzer in conducting
this test.

T. Collins
Operator Signature
TRAVIS COLLINS

Remarks: RFI test

Form 106-18000

Intoxilyzer Test Record and Checklist
ND0AG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer
North Dakota Model 8000
Location = BILL
06/23/2026

Alcohol Analyzer
SN 80-005958
8164.14.00 09/16
09:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:40
02 Std. Gas	0.078	09:41
03 Room Air	0.000	09:41
04 Std. Gas	0.078	09:41
05 Room Air	0.000	09:42
06 Std. Gas	0.078	09:42
07 Room Air	0.000	09:43

Lot No = 264437

Cyl No = 107

Exp Date = 01/29/2028

County = 04

Oper No. = 020533

T. Collins
Operator Signature
TRAVIS COLLINS

Remarks: ACA

Form 106-18000