

**INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT**

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number 80-005958	Instrument Location BILL
Reason for Install/Repair ☒ Install After Receiving From Crime Laboratory ☐ Install After Location Change ☐ Other (Specify) _____	

Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☐ 2. Telephone Line Connected to Intoxilyzer® 8000. *NA*
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☐ 7. Repair and/or Maintenance Performed (if any): _____
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☐ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency. *NA*
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>J. Collins</i>	Date 06/23/2026
--	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen <i>Anna Ingemansen</i>	Certified Date 07Jul2026
--	-----------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005958
Location = BILL 8164.14.00 09/16
06/23/2026 09:39

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#%&^*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 107
Standard Expiration? 01/29/2028
Oper No? 020533

Flow Cal. Date: 06/21/2013
Slope 716
Intercept -760436

IR Calibration Date: 04/07/2015
 3um 9um

0th Coef(*100):	-17179	-17694
1st Coef(*100):	259328	130890
2nd Coef(*100):	3473	1654
H2O adj(mg/l*10k):	425	422

***** Printer Test End *****



Operator Signature
TRAVIS COLLINS

Remarks:

Print Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005958
Location = BILL 8164.14.00 09:16
06/23/2026 09:49

Test	AC	Time
01 Diagnostic	OK	09:53
02 Room Air	0.000	09:53
03 *Subject Test	RFI*	09:53
04 Room Air	0.000	09:54

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST T
Sub DOB = 02/01/2000

Weight = 238
Cit = 1010010

Sub Sex = Male

Test = DUI

Dr. Lic. = ND/TES001234

Lot No = 264437

Cyl No = 107

Expiration Date = 01/29/2028

Oper No. = 020533

I followed the Approved Method
and the instructions displayed
by the Intoxilyzer in conducting
this test.

T. Collins
Operator Signature
TRAVIS COLLINS

Remarks: RFI test

Form 106-18000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005958
Location = BILL 8164.14.00 09:16
06/23/2026 09:40

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:40
02 Std. Gas	0.078	09:41
03 Room Air	0.000	09:41
04 Std. Gas	0.078	09:41
05 Room Air	0.000	09:42
06 Std. Gas	0.078	09:42
07 Room Air	0.000	09:43

Lot No = 264437

Cyl No = 107

Exp Date = 01/29/2028

County = 04

Oper No. = 020533

T. Collins
Operator Signature
TRAVIS COLLINS

Remarks: ACA

Form 106-18000