



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>William Earl</i>		
Location <i>DEVL</i>		Intoxilyzer® Serial Number <i>80-007096</i>
Gas Lot Number <i>25024080A3</i>	Gas Cylinder Number <i>04</i>	Gas Expiration Date <i>10/5/26</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. 0. *080* AC
B. 0. *079* AC
C. 0. *080* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>07/14/2026</i>
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen <i>Anna Ingemansen</i>	Date 17 July 2026


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007096
Location = DEVL 8164.16.00 09/18
07/11/2026 21:44

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:45
02 Std. Gas	0.080	21:45
03 Room Air	0.000	21:46
04 Std. Gas	0.079	21:46
05 Room Air	0.000	21:47
06 Std. Gas	0.080	21:47
07 Room Air	0.000	21:48

Lot No = 25024080A3
Cyl No = 04
Exp Date = 10/05/2026
County = 36 Oper No. = 132008



Operator Signature
WILLIAM EARL

Remarks:

Form 106-I8000