



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL



CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Joshua Gudvangen		
Location MILF		Intoxilyzer® Serial Number 80-007097
Gas Lot Number 264944	Gas Cylinder Number 313	Gas Expiration Date 05/12/2028

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. 0.0 71 AC
B. 0.0 71 AC
C. 0.0 71 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature  282	Date 07/17/2026
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen 	Date 29Jul2026


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = MILF 8164.16.00 09/18
07/17/2026 17:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:42
02 Std. Gas	0.079	17:42
03 Room Air	0.000	17:43
04 Std. Gas	0.079	17:43
05 Room Air	0.000	17:44
06 Std. Gas	0.079	17:44
07 Room Air	0.000	17:45

Lot No = 264944
Cyl No = 313
Exp Date = 05/12/2028
County = 51 Oper No. = 131334



Operator Signature
JOSHUA GUDVANGEN

Remarks:

Form 106-I8000