



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL


CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59281 (06/2018)

Serial Number 80-007097	Instrument Location MILF
Reason for Install/Repair ☒ Install After Receiving From Crime Laboratory ☐ Install After Location Change ☐ Other (Specify) _____	


Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☐ 7. Repair and/or Maintenance Performed (if any): _____
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature  282	Date 07/17/2026
---	--------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By Anna Ingemansen 	Certified Date 29Jul2026
---	-----------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = MILF 8164.16.00 09/18
07/17/2026 17:35

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264944
Standard Cyl #? 313
Standard Expiration? 05/12/2028
Oper No? 131334

Flow Cal. Date: 09/03/2020
Slope 660
Intercept -636264

IR Calibration Date: 03/24/2026
 3um 9um

0th Coef(*100):	-29472	-28763
1st Coef(*100):	250783	139555
2nd Coef(*100):	2701	1378
H2O adj(mg/l*10k):	627	553

***** Printer Test End *****

 212

Operator Signature
JOSHUA GUDVANGEN

Remarks:


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = MILF 8164.16.00 09/18
07/17/2026 17:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	17:42
02 Std. Gas	0.079	17:42
03 Room Air	0.000	17:43
04 Std. Gas	0.079	17:43
05 Room Air	0.000	17:44
06 Std. Gas	0.079	17:44
07 Room Air	0.000	17:45

Lot No = 264944
Cyl No = 313
Exp Date = 05/12/2028
County = 51 Oper No. = 131334


Operator Signature
JOSHUA GUDVANGEN

Remarks:

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-007097
Location = MILF 8164.16.00 09/18
07/17/2026 17:47

Test	AC	Time
01 Room Air	RFI*	17:48
02 Room Air	0.000	17:48

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 01/01/1990
Sub Sex = Male Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 264944
Cyl No = 313
Expiration Date = 05/12/2028
County = 51 Oper No. = 131334



Operator Signature
JOSHUA GUDVANGEN

Remarks:

Form 106-I8000