



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) MILES D RHONEMUS		
Location DEVL		Intoxilyzer® Serial Number 80-004944
Gas Lot Number 26-4437	Gas Cylinder Number 109	Gas Expiration Date 01/25/2028

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.080 AC
 - B. 0.080 AC
 - C. 0.080 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature 	Date 07/19/2026
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen	Date 21 July 2026


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004944
Location = DEVL 8164.14.00 09/16
07/19/2026 11:41

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	11:42
02 Std. Gas	0.080	11:43
03 Room Air	0.000	11:43
04 Std. Gas	0.080	11:44
05 Room Air	0.000	11:44
06 Std. Gas	0.080	11:45
07 Room Air	0.000	11:45

Lot No = 264437
Cyl No = 109
Exp Date = 01/29/2028
County = 36 Oper No. = 131613



Operator Signature
MILES RHONEMUS

Remarks:

Form 106-I8000