



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59281 (06/2018)

Serial Number <u>80-006673</u>	Instrument Location <u>BOTT</u>
Reason for Install/Repair ☐ Install After Receiving From Crime Laboratory ☐ Install After Location Change ☒ Other (Specify) <u>Annual Inspection at Crime Laboratory</u>	

Check When Done:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☒ 7. Repair and/or Maintenance Performed (if any): N/A
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN504096, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <u>Jeremy Arns</u>	Date <u>06/19/2026</u>
--	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By <u>Anna Ingemansen</u>	Certified Date <u>14 July 2026</u>
--	------------------------------------

This installation was performed at BOTT with the install being performed after the Annual Inspection at the Crime Laboratory. AEI 14Jul2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BOTT 8164.14.00 09/16
06/19/2026 11:51

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup


Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 264437
Standard Cyl #? 119
Standard Expiration? 01/29/2028
Oper No? 133044

Flow Cal. Date: 03/30/2017
Slope 697
Intercept -667725

IR Calibration Date: 04/15/2024
 3um 9um

0th Coef(*100):	-14904	-25220
1st Coef(*100):	290754	143378
2nd Coef(*100):	2231	1156
H2O adj(mg/l*10k):	289	324

***** Printer Test End *****



Operator Signature
JEREMY EVERS

Remarks:

Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BOTT 8164.14.00 09/16
06/19/2026 11:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:00
02 Std. Gas	0.078	12:00
03 Room Air	0.000	12:01
04 Std. Gas	0.077	12:01
05 Room Air	0.000	12:01
06 Std. Gas	0.077	12:02
07 Room Air	0.000	12:02

Lot No = 264437
Cyl No = 119
Exp Date = 01/29/2028
County = 05 Oper No. = 133044



Operator Signature
JEREMY EVERS

Remarks:

ACA Test

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006673
Location = BOTT 8164.14.00 09/16
06/19/2026 12:11

Test	AC	Time
01 Room Air	0.000	12:12
02 *Subject Test	RFI*	12:13
03 Room Air	0.000	12:13

*Invalid Test
Inhibited - RFI

Sub Name = RFI, TEST TEST
Sub DOB = 01/01/2001
Sub Sex = Unknown Weight = NA
Test = DUI Cit = NA
Dr. Lic. = ND/NA
Lot No = 264437
Cyl No = 119
Expiration Date = 01/29/2028
County = 05 Oper No. = 133044



Operator Signature
JEREMY EVERS

Remarks:

RFI TEST

Form 106-I8000