



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL



CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Kyle J. Johnson		
Location West Fargo PD	Intoxilyzer® Serial Number 80-006684	
Gas Lot Number 264944	Gas Cylinder Number 300	Gas Expiration Date 5/12/2028

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.081 AC
 - B. 0.081 AC
 - C. 0.081 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature  #423	Date 9/11/2026
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen 	Date 09/14/2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006684
Location = WFPD 8164.14.00 09/16
09/11/2026 15:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:39
02 Std. Gas	0.081	15:40
03 Room Air	0.000	15:40
04 Std. Gas	0.081	15:41
05 Room Air	0.000	15:41
06 Std. Gas	0.081	15:41
07 Room Air	0.000	15:42

Lot No = 264944
Cyl No = 300
Exp Date = 05/12/2028
County = 09 Oper No. = 130781

 #423
Operator Signature
KYLE JOHNSON

Remarks:

Form 106-I8000