

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION

INTOXILYZER® 8000 ANNUAL INSPECTION

Intoxilyzer® 8000 Serial Number: 80-00 6492

Inspection Location: TOXL

A. Pre-Inspection

1. Items with Instrument:

Gas Cylinder Yes or No (If Yes, Lot # 08425080A1 Cyl. # 3)

Keys Yes or No

Power Cord Yes or No

2. ☒ Download Data

3. ☒ Upload Operator File

4. ☒ Current Location Code: FYPD

5. ☒ Battery Check

Was the external battery pack replaced? Yes or No

6. ☒ O-Rings

Replaced Simulator O-Ring Yes or No

Replaced Breath Tube O-Ring Yes or No

B. General Setup and Checks:

1. ☒ Diagnostics passed and instrument in "Ready" mode

2. ☒ Breath tube heated

3. ☒ Date, time and location code (Level 2,E). Re-set if necessary.

Time Zone: CST or MDT (Time on test records will be in time zone circled).

4. ☒ Print test (Level 1,P). Sign and attach test record.

5. ☒ Tank monitor (Level 3,D,G).

Display: 1000 psi Regulator: 1000 psi

Display and Regulator $\pm$ 50 psi of each other Yes or No

Gas tank tare necessary? Yes or No

If Yes, display readings after tare (Level 3,M,C,G):

Display: _____ psi Regulator: _____ psi

C. Tests (Sign and attach test records):

1. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - Low AC (Level 1,C)

Known Value $\leq$ 0.03 AC: 0.020 AC

Sim. Ser #: MP5290

Lot #: 202501E

Exp. Date: 1/28/27

☒ Results $\pm$ 0.005 of known AC

2. ☒ Configure simulator for the following test (Level 1,S).

Wet Calibration Check - High AC (Level 1,C)

Known Value $\geq$ 0.25 AC: 0.300 AC

Sim. Ser #: MP3003

Lot #: 202408H

Exp. Date: 8/29/26

☒ Results $\pm$ 5% AC of known AC

3. ☒ Configure dry gas standard for the remaining tests (Level 1,S).

Known Value : 0.080 AC

Gas Cylinder Lot #: 264437

Cylinder #: 104

Exp. Date: 1/29/28

4. ☒ Interferent Check (Level 1,B)

Known Value: 0.10 AC + 0.05% Acetone

Sim. Ser #: DR7352

Lot #: 1CS9

Exp. Date: None

☒ Display reads "Interferent Detect"

5. ☒ Invalid Sample (Level 1,B)

☒ Display reads "Invalid Sample"

6. ☒ RFI Check (CMS Mode)

☒ Display reads "RFI Detect"

7. ☒ Dry Calibration Check (Level 1,C)

Test 1 0.080 Test 4 0.081 Test 7 0.081

Test 2 0.080 Test 5 0.080 Test 8 0.080

Test 3 0.080 Test 6 0.081 Test 9 0.080

Average 0.080

☒ Results $\pm$ 0.005 AC of known AC

D. Remarks/Maintenance: N/A

Instrument is acceptable to be used in the field. Yes or No

If No, state reason(s) why: _____

If Yes, change location code back to A.4 unless A.4 is TOXL. ☒

[Signature]
Inspector Signature

05 May 2026
Date

[Signature]
Reviewer

06 May 2026
Date

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:33

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 08425080A1
Standard Cyl #? 3
Standard Expiration? 05/05/2027
Oper No? 133237

Flow Cal. Date: 08/18/2015
Slope 669
Intercept -631515

IR Calibration Date: 08/18/2015

	3um	9um
0th Coef(*100):	-41685	-23145
1st Coef(*100):	269226	142194
2nd Coef(*100):	2671	1239
H2O adj(mg/l*10k):	872	437

***** Printer Test End *****


Operator Signature
ANNA INGEMANSEN

Remarks: Print Test

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:34

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:34
02 Std. Sol.	0.021	12:35
03 Room Air	0.000	12:36
04 Std. Sol.	0.021	12:36
05 Room Air	0.000	12:37
06 Std. Sol.	0.021	12:38
07 Room Air	0.000	12:38

08 Sim Temp = 34.0°C

Simul Ser No = MP5290

Std Sol No = 202501E

County = 08

Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks:

LOW AC Check
0.020AC

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501


CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:39

WET CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:40
02 Std. Sol.	0.299	12:41
03 Room Air	0.000	12:41
04 Std. Sol.	0.298	12:42
05 Room Air	0.000	12:43
06 Std. Sol.	0.298	12:43
07 Room Air	0.000	12:44

08 Sim Temp = 34.0°C

Simul Ser No = MP3003
Std Sol No = 202408H
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks:

High AC Check
0.300Ac

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:45

Test	AC	Time
01 Room Air	0.000	12:46
02 *Subject Test	INT*	12:46
03 Room Air	0.000	12:47

*Invalid Test
Interferent Detected

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INTERFERENT
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Interferent check
0.10Ac + 0.05% Acetone

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:51

Test	AC	Time
01 Room Air	0.000	12:52
02 Invalid Sample	X.XXX	12:52
03 Room Air	0.000	12:53

*Invalid Test - Mouth Alcohol

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = INVALID
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Invalid Sample
Mouth Alcohol

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:53

Test	AC	Time
01 Diagnostic	OK	12:54
02 Room Air	0.000	12:55
03 *Subject Test	RFI*	12:55
04 Room Air	0.000	12:55

*Invalid Test
Inhibited - RFI

Sub Name = TEST, DONOR2 NONE
Sub DOB = 07/25/1998
Sub Sex = Male Weight = NA
Test = OTH Cit = RFI
Dr. Lic. = ND/TES989643
Lot No = 264437
Cyl No = 104
Expiration Date = 01/29/2028
County = 08 Oper No. = 133237

I followed the Approved Method and the
instructions displayed by the Intoxilyzer
in conducting this test.



Operator Signature
ANNA INGEMANSEN

Remarks: RFI check

Form 106-I8000


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:56

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:56
02 Std. Gas	0.080	12:56
03 Room Air	0.000	12:57
04 Std. Gas	0.080	12:57
05 Room Air	0.000	12:58
06 Std. Gas	0.080	12:58
07 Room Air	0.000	12:59

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks:

Dry cal check
#1-3

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 12:59

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:00
02 Std. Gas	0.081	13:00
03 Room Air	0.000	13:01
04 Std. Gas	0.080	13:01
05 Room Air	0.000	13:01
06 Std. Gas	0.081	13:02
07 Room Air	0.000	13:02

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237



Operator Signature
ANNA INGEMANSEN

Remarks: Dry cal check
#4-6

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = TOXL 8164.14.00 09/16
05/05/2026 13:03

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	13:03
02 Std. Gas	0.081	13:03
03 Room Air	0.000	13:04
04 Std. Gas	0.080	13:04
05 Room Air	0.000	13:05
06 Std. Gas	0.080	13:05
07 Room Air	0.000	13:06

Lot No = 264437
Cyl No = 104
Exp Date = 01/29/2028
County = 08 Oper No. = 133237


Operator Signature
ANNA INGEMANSEN

Remarks:

Dry cal check
#7-9

Form 106-I8000