



ETHANOL GAS STANDARD CYLINDER REPORT
NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Monica Patneaud</i>		
Location <i>Standing Rock - Ft Yates, ND</i>		Intoxilyzer® Serial Number <i>80-006492</i>
Gas Lot Number <i>0842508A1</i>	Gas Cylinder Number <i>3</i>	Gas Expiration Date <i>5-5-2027</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
A. *0.082* AC
B. *0.079* AC
C. *0.080* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>10-1-26</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i> <i>Ann Ingemansen</i>	Date 01Oct2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = FYPD 8164.14.00 09/16
10/01/2026 14:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.080	14:41
03 Room Air	0.000	14:41
04 Std. Gas	0.079	14:42
05 Room Air	0.000	14:42
06 Std. Gas	0.080	14:43
07 Room Air	0.000	14:43

Lot No = 08425080A1
Cyl No = 3
Exp Date = 05/05/2027
County = 61 Oper No. = 130688


Operator Signature
MONICA PATNEAUD

Remarks:

Form 106-I8000