



INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION - TOXICOLOGY UNIT - BREATH ALCOHOL SECTION

SFN 59281 (05/2026)

Serial Number <i>80-006492</i>	Instrument Location <i>SRDC - Ft Yates</i>
-----------------------------------	---

REASON FOR INSTALL/REPAIR

- ☐ Install After Receiving From Crime Laboratory ☐ Install After Location Change
☒ Other - Specify: _____

CHECK WHEN DONE:

- ☒ 1. Surge Protector Installed/Property Grounded.
- ☒ 2. Telephone Line Connected to Intoxilyzer® 8000.
- ☒ 3. Breath Tube Heated.
- ☒ 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- ☒ 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 6. Run Tests:
 - ☒ A. Print Test (Level 1, Function P).
 - ☒ B. ACA Test (Level 1, Function C).
 - ☒ C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- ☐ 7. Repair and/or Maintenance Performed (if any) - Specify: _____
- ☒ 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature <i>[Signature]</i>	Date <i>10/1/26</i>
---	------------------------

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By (print name) Anna Ingemansen	Reviewed/Certified By Signature <i>Anna Ingemansen</i>	Certified Date 01Oct2026
---	---	-----------------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = FYPD 8164.14.00 09/16
10/01/2026 14:35

***** Printer Test *****

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=_|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#\$\$%^&*()_+?

Current Instrument Setup

Data Entry Mode: Enabled
Start Test Sequence: DABACABA
Display Prelim Rslt? Yes
Display Third Digit? Yes
Inhib Printer(Y/N)? No
Display Volume? No
Disable On Memfull? Yes
of Print Copies? 1
Select Std (D/W/I)? Dry
Standard Value? 0.080
Standard Lot #? 08425080A1
Standard Cyl #? 3
Standard Expiration? 05/05/2027
Oper No? 130688

Flow Cal. Date: 08/18/2015
Slope 669
Intercept -631515

IR Calibration Date: 08/18/2015
 3um 9um

0th Coef(*100):	-41685	-23145
1st Coef(*100):	269226	142194
2nd Coef(*100):	2671	1239
H2O adj(mg/l*10k):	872	437

***** Printer Test End *****



Operator Signature
MONICA PATNEAUD

Remarks:

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = FYPD 8164.14.00 09/16
10/01/2026 14:39

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	14:40
02 Std. Gas	0.080	14:41
03 Room Air	0.000	14:41
04 Std. Gas	0.079	14:42
05 Room Air	0.000	14:42
06 Std. Gas	0.080	14:43
07 Room Air	0.000	14:43

Lot No = 08425080A1
Cyl No = 3
Exp Date = 05/05/2027
County = 61 Oper No. = 130688



Operator Signature
MONICA PATNEAUD

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006492
Location = FYPD 8164.14.00 09/16
10/01/2026 14:47

Test	AC	Time
01 Room Air	0.000	14:49
02 Reference	RFI*	14:49
03 Room Air	0.000	14:50

*Invalid Test
Inhibited - RFI

Sub Name = TEST, TEST TEST
Sub DOB = 09/09/1999
Sub Sex = Male Weight = 190
Test = DUI Cit = NA
Dr. Lic. = ND/ND
Lot No = 08425080A1
Cyl No = 3
Expiration Date = 05/05/2027
County = 61 Oper No. = 130688



Operator Signature
MONICA PATNEAUD

Remarks:

Form 106-I8000