



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) Allen, Kathryn			090801
Location Cass County Jail - Booking		Intoxilyzer® Serial Number 80-064954	
Gas Lot Number 26-4437	Gas Cylinder Number 112	Gas Expiration Date 01/29/2028	

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0.079 AC
 - B. 0.080 AC
 - C. 0.079 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature [Signature] #3671	Date 10/02/20
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen	Date 07Oct2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004954
Location = CASS 8164.14.00 09/16
10/02/2026 15:47

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	15:48
02 Std. Gas	0.079	15:48
03 Room Air	0.000	15:48
04 Std. Gas	0.080	15:49
05 Room Air	0.000	15:49
06 Std. Gas	0.079	15:50
07 Room Air	0.000	15:50

Lot No = 264437
Cyl No = 112
Exp Date = 01/29/2028
County = 09 Oper No. = 090801


Operator Signature
KATHRYN ALLEN

Remarks:

Good Test !!
😊

Form 106-I8000