



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (05/2026)

Chemical Test Operator Name (Print) <i>Gross, Danny Joe</i>		
Location <i>Lisbon PD</i>	Intoxilyzer® Serial Number <i>80-006512</i>	
Gas Lot Number <i>264944</i>	Gas Cylinder Number <i>309</i>	Gas Expiration Date <i>5/12/2028</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. 0. *0.080* AC
 - B. 0. *0.081* AC
 - C. 0. *0.081* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>10/03/2026</i>
---	------------------------

Crime Laboratory Use Only

Reviewed By (Print Name) Anna Ingemansen	Reviewed By Signature <i>Anna Ingemansen</i>	Date 05Oct2026
---	--	-------------------

During the first ACA test, it was determined that the gas lot number was incorrect (2264944). Anna Ingemansen reached out and Jonathan Oien updated the lot number and sent in another ACA test (Last page) to show the lot number had been updated to the correct lot number of 264944. This does not affect the installation of the ethanol gas standard cylinder.
AEI 10/5/2026


Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006513
Location = LISB 8164.14.00 09/16
10/03/2026 18:07

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:08
02 Std. Gas	0.080	18:09
03 Room Air	0.000	18:09
04 Std. Gas	0.081	18:09
05 Room Air	0.000	18:10
06 Std. Gas	0.081	18:10
07 Room Air	0.000	18:11

Lot No = 2264944
Cyl No = 309
Exp Date = 05/12/2028
County = 37 Oper No. = 133737



Operator Signature
DANNY GROSS

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006513
Location = LISB 8164.14.00 09/16
10/05/2026 10:43

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:44
02 Std. Gas	0.080	10:44
03 Room Air	0.000	10:45
04 Std. Gas	0.081	10:45
05 Room Air	0.000	10:46
06 Std. Gas	0.081	10:46
07 Room Air	0.000	10:47

Lot No = 264944
Cyl No = 309
Exp Date = 05/12/2028
County = 37 Oper No. = 131629


Operator Signature
JONATHAN OIEN

Remarks:

I certify that the foregoing
is a true & correct copy of the
original document on file in this office.

Form 106-I8000