



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Logan Mellman</i>		
Location <i>Mandan</i>		Intoxilyzer® Serial Number <i>80-006506</i>
Gas Lot Number <i>564944</i>	Gas Cylinder Number <i>504</i>	Gas Expiration Date <i>5-12-2028</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.080* AC
 - B. *0.080* AC
 - C. *0.080* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>10-7-2026</i>
Reviewed By (Crime Laboratory Use Only) Anna Ingemansen <i>Ann Ingemansen</i>	Date 08Oct2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006506
Location = MAND 8164.14.00 09/16
10/07/2026 18:16

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	18:16
02 Std. Gas	0.080	18:17
03 Room Air	0.000	18:17
04 Std. Gas	0.080	18:18
05 Room Air	0.000	18:18
06 Std. Gas	0.080	18:18
07 Room Air	0.000	18:19

Lot No = 264944
Cyl No = 304
Exp Date = 05/12/2028
County = 30 Oper No. = 133572


Operator Signature
LOGAN MOLLMAN

Remarks: *Install*

Form 106-I8000