



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (08/2018)

Chemical Test Operator Name (Print) <i>Demetrius Jamerson</i>		
Location <i>NEWT (TAT)</i>		Intoxilyzer® Serial Number <i>80-006691</i>
Gas Lot Number <i>264944</i>	Gas Cylinder Number <i>294</i>	Gas Expiration Date <i>5/12/2028</i>

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☐ 2. Perform an ACA Test (Level 1, Function C).
Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
 - A. *0.077* AC
 - B. *0.076* AC
 - C. *0.077* AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
 - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
 - B. ACA Test Record.

Chemical Test Operator Signature <i>[Signature]</i>	Date <i>10-7-26</i>
Reviewed By (Crime Laboratory Use Only) <i>Anna Ingemansen</i>	Date <i>07Oct2026</i>

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006691
Location = NEWT 8164.14.00 09/16
10/07/2026 12:50

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	12:51
02 Std. Gas	0.077	12:51
03 Room Air	0.000	12:52
04 Std. Gas	0.076	12:52
05 Room Air	0.000	12:53
06 Std. Gas	0.077	12:53
07 Room Air	0.000	12:54

Lot No = 264944

Cyl No = 294

Exp Date = 05/12/2028

County = 60

Oper No. = 132358



Operator Signature
DEMETRIUS JAMERSON

Remarks:

New Cylinder install

Form 106-I8000