



ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL

CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM

SFN 59282 (05/2026)

Chemical Test Operator Name (Print) <i>Beth Isaak</i>		
Location <i>Killdeer PD</i>	Intoxilyzer® Serial Number <i>80-004194</i>	
Gas Lot Number <i>264944</i>	Gas Cylinder Number <i>315</i>	Gas Expiration Date <i>5-12-2028</i>

Check When Done:

☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).

☒ 2. Perform an ACA Test (Level 1, Function C).

Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

A. 0. *079* AC

B. 0. *079* AC

C. 0. *078* AC

☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.

☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.

☒ 5. Send the Following to the Crime Laboratory:

A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).

B. ACA Test Record.

Chemical Test Operator Signature <i>Beth Isaak</i>	Date <i>10-9-26</i>
--	---------------------

Crime Laboratory Use Only

Reviewed By (Print Name) Anna Ingemansen	Reviewed By Signature <i>Anna Ingemansen</i>	Date 09Oct2026
---	---	-------------------

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-004194
Location = KDPD 8164.14.00 09/16
10/09/2026 10:34

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	10:35
02 Std. Gas	0.079	10:35
03 Room Air	0.000	10:36
04 Std. Gas	0.079	10:36
05 Room Air	0.000	10:37
06 Std. Gas	0.078	10:37
07 Room Air	0.000	10:37

Lot No = 264944
Cyl No = 315
Exp Date = 05/12/2028
County = 13 Oper No. = 131919


Operator Signature
BETH ISAAK

Remarks:

Form 106-I8000