ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (08/2018)

Chemical Test Operator Name (Print): Howard, Austin L		
Location: Powers Lake police Department	Intoxilyzer® Serial Number: 80-006687	
Gas Lot Number: ~~302~~ ALT 264944	Gas Cylinder Number: ~~364944~~ ALT 302	Gas Expiration Date: 5/12/28

Check When Done:

- ☒ 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- ☒ 2. Perform an ACA Test (Level 1, Function C).
  Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).
  - A. 0.079 AC
  - B. 0.078 AC
  - C. 0.078 AC
- ☒ 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- ☒ 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- ☒ 5. Send the Following to the Crime Laboratory:
  - A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  - B. ACA Test Record.

Chemical Test Operator Signature: Austin Howard	Date: 10/22/26
Reviewed By (Crime Laboratory Use Only): Anna Ingemansen Anna Ingemansen	Date: 11/02/2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-006687
Location = PLPD 8164.14.00 09/16
10/22/2026 21:53

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	21:54
02 Std. Gas	0.079	21:54
03 Room Air	0.000	21:55
04 Std. Gas	0.078	21:55
05 Room Air	0.000	21:55
06 Std. Gas	0.078	21:56
07 Room Air	0.000	21:56

Lot No = 264944
Cyl No = 302
Exp Date = 05/12/2028
County = 07 Oper No. = 132545



Operator Signature
AUSTIN HOWARD

Remarks:

Form 106-I8000