ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (05/2026)

Chemical Test Operator Name (Print)		
Daniel Magelky		
Location: BMDC	Intoxilyzer® Serial Number: 80-005358	
Gas Lot Number: 264437	Gas Cylinder Number: 113	Gas Expiration Date: 1/29/2028

Check When Done:

- [X] 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- [X] 2. Perform an ACA Test (Level 1, Function C).

  Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

  A. 0.081 AC
  B. 0.081 AC
  C. 0.081 AC
- [X] 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- [X] 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- [X] 5. Send the Following to the Crime Laboratory:
  A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  B. ACA Test Record.

Chemical Test Operator Signature	Date
Dal Mag	11/5/26

Crime Laboratory Use Only

Reviewed By (Print Name)	Reviewed By Signature	Date
Anna Ingemansen	Anna Ingemansen	17Nov2026

FORM 105-G

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005358
Location = BMDC — 8164.14.00 09/16
11/05/2026 — 09:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Gas	0.081	09:22
03 Room Air	0.000	09:23
04 Std. Gas	0.081	09:23
05 Room Air	0.000	09:24
06 Std. Gas	0.081	09:24
07 Room Air	0.000	09:25

Lot No = 264437
Cyl No = 113
Exp Date = 01/29/2028
County = 08 — Oper No. = 133598

Operator Signature
DANIEL MAGELKY

Remarks:

Form 106-I8000