INTOXILYZER® 8000 INSTALLATION AND REPAIR CHECKOUT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION - TOXICOLOGY UNIT - BREATH ALCOHOL SECTION
SFN 59281 (05/2026)

Serial Number	Instrument Location
80-005358	BMDC

REASON FOR INSTALL/REPAIR

- [X] Install After Receiving From Crime Laboratory
- [ ] Install After Location Change
- [ ] Other - Specify: _____

CHECK WHEN DONE:

- [X] 1. Surge Protector Installed/Property Grounded.
- [X] 2. Telephone Line Connected to Intoxilyzer® 8000.
- [X] 3. Breath Tube Heated.
- [X] 4. Enter Preliminary Data (i.e. Date, Time, DST (Y), and Location; Level 2, Function E).
- [X] 5. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- [X] 6. Run Tests:
  - [X] A. Print Test (Level 1, Function P).
  - [X] B. ACA Test (Level 1, Function C).
  - [X] C. Radio Frequency Interference (RFI) Test (CMS Mode or Level 1, Function B or C; Key Radio During Test).
- [ ] 7. Repair and/or Maintenance Performed (if any) - Specify: _____
- [X] 8. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- [X] 9. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- [X] 10. Send the Following to the Crime Laboratory: Completed Intoxilyzer® 8000 Installation and Repair Checkout (SFN59281, Form 104-G), Print Test, ACA Test, and RFI Test.

Field Inspector Signature	Date
[signature]	11/5/26

Crime Laboratory Use Only

This installation has been reviewed and the instrument is approved to be used for the analysis of breath to determine alcohol concentration from the date the Field Inspector performed the installation. This record on file at the Office of Attorney General, Crime Laboratory Division, in the County of Burleigh, North Dakota, is certified to be a true and correct copy of the documents received.

Reviewed/Certified By (print name)	Reviewed/Certified By Signature	Certified Date
Anna Ingemansen	[signature]	17Nov2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer Alcohol Analyzer
North Dakota Model 8000 SN 80-005358
Location = BMDC 8164.14.00 09/16
11/05/2026 09:17

********** Printer Test **********

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*()_+?

abcdefghijklmnopqrstuvwxyz1234567890-=|
ABCDEFGHIJKLMNOPQRSTUVWXYZ!@#$%^&*()_+?

Current Instrument Setup

Data Entry Mode:	Enabled
Start Test Sequence:	DABACABA
Display Prelim Rslt?	Yes
Display Third Digit?	Yes
Inhib Printer(Y/N)?	No
Display Volume?	No
Disable On Memfull?	Yes
# of Print Copies?	1
Select Std (D/W/I)?	Dry
Standard Value?	0.080
Standard Lot #?	264437
Standard Cyl #?	113
Standard Expiration?	01/29/2028
Oper No?	133598
Flow Cal. Date:	06/04/2019
Slope	570
Intercept	4079
IR Calibration Date:	06/22/2023

	3um	9um
0th Coef(*100):	-5061	-17514
1st Coef(*100):	262541	141786
2nd Coef(*100):	4968	1592
H2O adj(mg/l*10k):	286	373

********** Printer Test End **********

Operator Signature
DANIEL MAGELKY

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005358
Location = BMDC — 8164.14.00 09/16
11/05/2026 — 09:21

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:22
02 Std. Gas	0.081	09:22
03 Room Air	0.000	09:23
04 Std. Gas	0.081	09:23
05 Room Air	0.000	09:24
06 Std. Gas	0.081	09:24
07 Room Air	0.000	09:25

Lot No = 264437
Cyl No = 113
Exp Date = 01/29/2028
County = 08 — Oper No. = 133598

Operator Signature
DANIEL MAGELKY

Remarks:

Form 106-I8000

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005358
Location = BMDC — 8164.14.00 09/16
11/05/2026 — 09:17

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	09:18
02 Std. Gas	0.080	09:18
03 Room Air	0.000	09:19
04 Std. Gas	0.080	09:19
05 Room Air	RFI*	09:19
06 Room Air	0.000	09:20

*Invalid Test
Inhibited - RFI

Lot No = 264437
Cyl No = 113
Exp Date = 01/29/2028
County = 08 — Oper No. = 133598

Operator Signature
DANIEL MAGELKY

Remarks:

Form 106-I8000