ETHANOL GAS STANDARD CYLINDER REPORT

NORTH DAKOTA OFFICE OF ATTORNEY GENERAL
CRIME LABORATORY DIVISION-TOXICOLOGY SECTION/BREATH ALCOHOL PROGRAM
SFN 59282 (05/2026)

Chemical Test Operator Name (Print)		
Pamela Pehl		
Location: Stanton	Intoxilyzer® Serial Number: 80-005944	
Gas Lot Number: 264944	Gas Cylinder Number: 305	Gas Expiration Date: 05/12/2028

Check When Done:

- [x] 1. Scan/Enter Gas Cylinder Information (Level 1, Function S).
- [x] 2. Perform an ACA Test (Level 1, Function C).

  Write in the ACA Test Results (Reported to 3 Digits, Ex. 0.081).

  A. 0.080 AC
  B. 0.079 AC
  C. 0.079 AC
- [x] 3. Complete the Top Portion of the Intoxilyzer® Record (SFN50496, Form 120-G) and Place it by the Intoxilyzer® for Use.
- [x] 4. File Previous Intoxilyzer® Record (SFN50496, Form 120-G) at the Intoxilyzer® Location at the Agency.
- [x] 5. Send the Following to the Crime Laboratory:
  A. Ethanol Gas Standard Cylinder Report (SFN59282, Form 105-G).
  B. ACA Test Record.

Chemical Test Operator Signature	Date
P Pehl	11/16/26

Crime Laboratory Use Only

Reviewed By (Print Name)	Reviewed By Signature	Date
Anna Ingemansen	Anna Ingemansen	17Nov2026

Intoxilyzer Test Record and Checklist
NDOAG Crime Lab. Div., Bismarck, ND 58501

CMI, Inc. Intoxilyzer — Alcohol Analyzer
North Dakota Model 8000 — SN 80-005944
Location = STAT — 8164.14.00 09/16
11/16/2026 — 16:02

DRY CAL CHECK

Test	AC	Time
01 Room Air	0.000	16:02
02 Std. Gas	0.080	16:03
03 Room Air	0.000	16:03
04 Std. Gas	0.079	16:04
05 Room Air	0.000	16:04
06 Std. Gas	0.079	16:05
07 Room Air	0.000	16:05

Lot No = 264944
Cyl No = 305
Exp Date = 05/12/2028
County = 29 — Oper No. = 120906



Operator Signature
PAMELA PEHL

Remarks:

Form 106-I8000